

WHO International Standard
WHO 1st Reference Reagent for Lentiviral Vector Integration Site
Analysis
NIBSC code: 18/144

Instructions for use (Version 3.0, Dated 30/09/2020)

1. INTENDED USE

The Reference Reagent was established in 2019 by the Expert Committee on Biological Standardization of the World Health Organization (WHO) as the WHO 1st International Reference Reagent (RR) for lentiviral vector (LV) integration site analysis.

The coded ampoule (NIBSC Code 18/144) contains freeze-dried, purified genomic DNA extracted from human cells. The material was tested by external laboratories and showed suitability as a qualitative Reference Reagent for LV integration site analysis, with a confident detection of the ten defined LV integration sites (Table 1).

This Reference Reagent should not be used for any other purpose. Data analysis must be focussed on lentiviral vector integration.

2. CAUTION

This preparation is not for administration to humans or animals in the human food chain.

The preparation contains material of human origin, and either the final product or the source materials, from which it is derived, have been tested and found negative for HBsAg, anti-HIV and HCV RNA. As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols. Care should be exercised in opening ampoules or vials, to avoid cuts.

3. UNITAGE

The Reference Reagent was tested in an International collaborative study involving thirty-one laboratories from thirteen countries. The consensus ten LV integration sites were obtained from the most concordant results. Not all intergration sites have been necessarily reported. The number of integration sites is provided as a reference value for information and NOT as a formally assigned unit.

Table 1. LV Integration sites

Chromosome	Start (approx.)	Gene
21	43867986	AGPAT3
6	34299319	NUDT3
3	50145157	SEMA3F
4	93625425	GRID2
8	39055454	ADAM9
12	57274964	R3HDM2
9	163465	CBWD1
7	114508378	FOXP2
22	39825823	ENTHD1
17	1821489	SMYD4

4. CONTENTS

Country of origin of biological material: United Kingdom.

The Reference Reagent contains approximately 5µg freeze-dried and purified genomic DNA extracted from human cell lines. The genomic

DNAs were extracted using a "salting out" method and diluted in Tris-EDTA buffer with 5mg/mL Trehalose before freeze-drying.

5. STORAGE

Store all unopened ampoules of freeze-dried materials at -20°C or below.

Please note: because of the inherent stability of lyophilized material, NIBSC may ship these materials at ambient temperature.

6. DIRECTIONS FOR OPENING

DIN ampoules have an 'easy-open' coloured stress point, where the narrow ampoule stem joins the wider ampoule body. Various types of ampoule breaker are available commercially. To open the ampoule, tap the ampoule gently to collect material at the bottom (labelled) end and follow manufactures instructions provided with the ampoule breaker.

7. USE OF MATERIAL

No attempt should be made to weigh out any portion of the freeze-dried material prior to reconstitution

- a. Open the ampoule as described in section 6, above.
- b. Reconstitute the freeze-dried material at room temperature with e.g. nuclease-free water.
- c. Allow the materials to reconstitute for 1 hour at room temperature and then pipette well to \mbox{mix} .
- d. Transfer the sample to a nuclease-free tube using a pipette, ensuring the maximum available volume is collected.
- e. After being reconstituted in, e.g. $200\mu L$ /ampoule, nuclease-free water, the gDNA concentration of the material will be approximately $25 ng/\mu L$ based on QuBit quantitation in Tris-EDTA buffer (10mM tris, 1mMEDTA) containing 5 mg/mL Trehalose. The possible appearance of white flecks in the materials should not be of concern.
- f. To be able to detect the ten defined LV integration sites in order to validate end-users' study protocols for LV integration site analysis.

8. STABILITY

Materials are held at NIBSC within assured, temperature-controlled storage facilities. Materials should be stored on receipt as indicated on the label.

NIBSC follows the policy of WHO with respect to its reference standards. It is the policy of the WHO to not assign an expiry date to either international reference reagents. They remain valid with the assigned values and status until withdrawn or amended.

Accelerated degradation studies have indicated that these materials are suitably stable when stored at -20°C or below, for the assigned values to remain valid until the materials are withdrawn or replaced. These studies have also shown that the materials are suitably stable for shipment at ambient temperature without any effect on the assigned values.

It is highly recommended that the materials is used on the day it is reconstituted and is not stored. However, in-house analysis determined reconstituted freeze-dried genomic DNA to be stable for up to 1 week at +4°C (or one month at -20°C). Care should be taken to avoid cross-contamination with other samples.

Users who have any data supporting any deterioration in the characteristics of materials are encouraged to contact NIBSC.





9. REFERENCES

WHO document WHO/BS/2019.xxx

10. ACKNOWLEDGEMENTS

We would like to thank Professor Didier Trono (EPFL, Lausanne) for his donation of the Lentiviral plasmids to make the project possible. We greatly appreciate the significant contributions of all collaborative study participants.

11. FURTHER INFORMATION

Further information can be obtained as follows;

This material: enquiries@nibsc.org

WHO Biological Standards:

http://www.who.int/biologicals/en/

JCTLM Higher order reference materials:

http://www.bipm.org/en/committees/jc/jctlm/

Derivation of International Units:

http://www.nibsc.org/standardisation/international_standards.aspx

Ordering standards from NIBSC:

http://www.nibsc.org/products/ordering.aspx

NIBSC Terms & Conditions:

http://www.nibsc.org/terms_and_conditions.aspx

12. CUSTOMER FEEDBACK

Customers are encouraged to provide feedback on the suitability or use of the material provided or other aspects of our service. Please send any comments to enquiries@nibsc.org

13. CITATION

biological waste.

In all publications, including data sheets, in which this material is referenced, it is important that the preparation's title, its status, the NIBSC code number, and the name and address of NIBSC are cited and cited correctly.

14. MATERIAL SAFETY SHEET

Classification in accordance with Directive 2000/54/EC, Regulation (EC) No 1272/2008: Not applicable or not classified

(LO) NO 1272/2000. Not applicable of flot classified					
Physical and Chemical properties					
Physical appearance:		Corrosive:	No		
White cake					
Stable: Yes		Oxidising:	No		
Hygroscopic: No		Irritant:	Unknown		
Flammable: No		Handling:See caution, Section 2			
Other (specify): N/A					
Toxicological properties					
Effects of inhalation: Not		established, avoid inhalation			
		established, avoid ingestion			
Effects of skin absorption: Not		established, avoid contact with skin			
Suggested First Aid					
Inhalation: Seek medical advice					
Ingestion: Seek	Seek medical advice				
	Wash with copious amounts of water. Seek				
medical advice					
Contact with skin: Wash	ontact with skin: Wash thoroughly with water.				
Action on Spillage and Method of Disposal					
Spillage of ampoule contents should be taken up with absorbent					
material wetted with an appropriate disinfectant. Rinse area with an appropriate disinfectant followed by water.					

National Institute for Biological Standards and Control

Absorbent materials used to treat spillage should be treated as

Potters Bar, Hertfordshire, EN6 3QG. T +44 (0)1707 641000, nibsc.org WHO International Laboratory for Biological Standards, UK Official Medicines Control Laboratory

15. LIABILITY AND LOSS

In the event that this document is translated into another language, the English language version shall prevail in the event of any inconsistencies between the documents.

Unless expressly stated otherwise by NIBSC, NIBSC's Standard Terms and Conditions for the Supply of Materials (available at http://www.nibsc.org/About_Us/Terms_and_Conditions.aspx or upon request by the Recipient) ("Conditions") apply to the exclusion of all other terms and are hereby incorporated into this document by reference. The Recipient's attention is drawn in particular to the provisions of clause 11 of the Conditions.

16. INFORMATION FOR CUSTOMS USE ONLY

Country of origin for customs purposes*: United Kingdom

* Defined as the country where the goods have been produced and/or sufficiently processed to be classed as originating from the country of supply, for example a change of state such as freeze-drying.

Net weight: 0.003g

Toxicity Statement: Toxicity not assessed

Veterinary certificate or other statement if applicable.

Attached: No

17. CERTIFICATE OF ANALYSIS

NIBSC does not provide a Certificate of Analysis for WHO Biological Reference Materials because they are internationally recognised primary reference materials fully described in the instructions for use. The reference materials are established according to the WHO Recommendations for the preparation, characterization and establishment of international and other biological reference standards http://www.who.int/bloodproducts/publications/TRS932Annex2_Inter_biol efstandardsrev2004.pdf (revised 2004). They are officially endorsed by the WHO Expert Committee on Biological Standardization (ECBS) based on the report of the international collaborative study which established their suitability for the intended use.

