

WHO International Standard
INTERFERON ALPHA n1, (HUMAN, lymphoblastoid cell-derived)
NIBSC code: 95/568
Instructions for use
(Version 5.0, Dated 16/04/2013)

1. INTENDED USE

This material is the 2nd WHO International Standard for human interferon alpha of lymphoblastoid cell origin. This material replaces the 1st International Standard for lymphoblastoid interferon coded Ga23-901-532. It is intended for use as the primary biological reference standard in bioassays for interferon alpha n1 (human interferon alpha, lymphoblastoid cell-derived).

2. CAUTION

This preparation is not for administration to humans or animals in the human food chain.

The preparation contains material of human origin, and either the final product or the source materials, from which it is derived, have been tested and found negative for HBsAg, anti-HIV and HCV RNA. As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols. Care should be exercised in opening ampoules or vials, to avoid cuts.

3. UNITAGE

38, 000 International Units per ampoule.

4. CONTENTS

Country of origin of biological material: United Kingdom. Each ampoule contains a freeze-dried residue comprising, under an atmosphere of nitrogen:

Interferon alpha n1, approximately 250 ng 6-salt phosphate buffered saline pH 7.0 6.0 mg human serum albumin

The interferon alpha n1 was derived from human cells.

5. STORAGE

For economy of use, it is recommended that the final solution be subdivided into several small aliquots and stored at -40 $^{\circ}$ C or below. Avoid repeated thawing/freezing. Unopened ampoules should be stored at -20 $^{\circ}$ C.

Please note: because of the inherent stability of lyophilized material, NIBSC may ship these materials at ambient temperature.

6. DIRECTIONS FOR OPENING

DIN ampoules have an 'easy-open' coloured stress point, where the narrow ampoule stem joins the wider ampoule body. Various types of ampoule breaker are available commercially. To open the ampoule, tap the ampoule gently to collect material at the bottom (labelled) end and follow manufactures instructions provided with the ampoule breaker.

7. USE OF MATERIAL

No attempt should be made to weigh out any portion of the freeze-dried material prior to reconstitution

No attempt should be made to weigh out any portion of the freeze-dried material. Dissolve the total contents of the ampoule in 0.5ml of sterile distilled water and transfer to a sterile container. Rinse the ampoule with about 0.4ml of sterile distilled water and add to the first solution. Make up the total volume to 1.0ml with sterile distilled water. The final solution will contain IFN alpha n1 at a concentration of 38000 International Units per

ml. Use carrier protein where dilution is required. It is recommended that initial dilutions, i.e. 1:10, 1:100, are either made in cell culture medium containing - 5%v/v -10%v/v calf serum or in phosphate-buffered saline, pH 7.0-7.4, containing 0.3%v/v bovine casein to prevent adsorption of IFN to container surfaces.

8. STABILITY

Reference materials are held at NIBSC within assured, temperature-controlled storage facilities and they should be stored on receipt as indicated on the label. It is the policy of WHO not to assign an expiry date to their international reference materials. Accelerated degradation studies have indicated that this material is suitably stable, when stored at -20°C or below, for the assigned values to remain valid until the material is withdrawn or replaced. These studies have also shown that the material is suitably stable for shipment at ambient temperature without any effect on the assigned values. Once reconstituted, diluted or aliquoted, users should determine the stability of the material according to their own method of preparation, storage and use. Users who have data supporting any deterioration in the characteristics of any reference preparation are encouraged to contact NIBSC.

9. REFERENCES

Meager, A, Gaines Das, R, Zoon K. and Mire-Sluis, A. (2001) Establishment of new and replacement World Health Organisation International Biological Standards for human interferon alpha and omega. Journal of Immunological Methods, 257, 17-33.

This standard was produced under WHO guidelines as cited in the WHO Technical Reports Series 800, 1990, Annex 4.

10. ACKNOWLEDGEMENTS

N/A

11. FURTHER INFORMATION

Further information can be obtained as follows; This material: enquiries@nibsc.org WHO Biological Standards: http://www.who.int/biologicals/en/
JCTLM Higher order reference materials: http://www.bipm.org/en/committees/jc/jctlm/
Derivation of International Units:

http://www.nibsc.org/standardisation/international_standards.aspx

Ordering standards from NIBSC:

http://www.nibsc.org/products/ordering.aspx

NIBSC Terms & Conditions:

http://www.nibsc.org/terms_and_conditions.aspx

12. CUSTOMER FEEDBACK

Customers are encouraged to provide feedback on the suitability or use of the material provided or other aspects of our service. Please send any comments to enquiries@nibsc.org

13. CITATION

In all publications, including data sheets, in which this material is referenced, it is important that the preparation's title, its status, the NIBSC code number, and the name and address of NIBSC are cited and cited correctly.

14. MATERIAL SAFETY SHEET



Potters Bar, Hertfordshire, EN6 3QG. T +44 (0)1707 641000, nibsc.org WHO International Laboratory for Biological Standards, UK Official Medicines Control Laboratory





Classification in accordance with Directive 2000/54/EC, Regulation (EC) No 1272/2008: Not applicable or not classified

Physical and Chemical properties		
Physical	Corrosive:	No
appearance:		
Freeze-dried white		
powder	Out distance	NI -
Stable: Yes	Oxidising:	No
Hygroscopic: Yes	Irritant:	No
Flammable:	Handling:	See caution, Section 2
No		
Other (specify): Contains material of human origin		
Toxicological properties		
Effects of inhalation: Not established, avoid inhalation		
Effects of ingestion: Not established, avoid ingestion		
Effects of skin absorption: Not established, avoid contact with skin		
Suggested First Aid		
Inhalation: Seek medical advice		
Ingestion: Seek medical advice		
Contact with eyes: Wash with copious amounts of water. Seek medical advice		
Contact with skin: W	Vash thoroughly with water.	
Action on Spillage and Method of Disposal		
Spillage of ampoule contents should be taken up with absorbent		

http://www.who.int/bloodproducts/publications/TRS932Annex2_Inter_biol efstandardsrev2004.pdf (revised 2004). They are officially endorsed by the WHO Expert Committee on Biological Standardization (ECBS) based on the report of the international collaborative study which established their suitability for the intended use.

15. LIABILITY AND LOSS

biological waste.

appropriate disinfectant followed by water.

In the event that this document is translated into another language, the English language version shall prevail in the event of any inconsistencies between the documents.

material wetted with an appropriate disinfectant. Rinse area with an

Absorbent materials used to treat spillage should be treated as

Unless expressly stated otherwise by NIBSC, NIBSC's Standard Terms and Conditions for the Supply of Materials (available at http://www.nibsc.org/About_Us/Terms_and_Conditions.aspx or upon request by the Recipient) ("Conditions") apply to the exclusion of all other terms and are hereby incorporated into this document by reference. The Recipient's attention is drawn in particular to the provisions of clause 11 of the Conditions.

16. INFORMATION FOR CUSTOMS USE ONLY

Country of origin for customs purposes*: United Kingdom
* Defined as the country where the goods have been produced and/or
sufficiently processed to be classed as originating from the country of
supply, for example a change of state such as freeze-drying.

Net weight: 1g

Toxicity Statement: Toxicity not assessed

Veterinary certificate or other statement if applicable.

Attached: No

17. CERTIFICATE OF ANALYSIS

NIBSC does not provide a Certificate of Analysis for WHO Biological Reference Materials because they are internationally recognised primary reference materials fully described in the instructions for use. The reference materials are established according to the WHO Recommendations for the preparation, characterization and establishment of international and other biological reference standards

National Institute for Biological Standards and Control,

Potters Bar, Hertfordshire, EN6 3QG. T +44 (0)1707 641000, nibsc.org WHO International Laboratory for Biological Standards, UK Official Medicines Control Laboratory

